


ZOLL ECG Electrodes
Ag/AgCl Pregelled Monitoring Electrodes

Contains
30

ZOLL

Z9526-93 Rev. D



Cautions:

- Do not use if gel is dry.
- Replacement of electrodes after each 24 hour period is suggested. Vary placement to minimize skin sensitivity.

ZOLL
Advancing Resuscitation. Today®



ZOLL Medical Corporation
269 Mill Road
Chelmsford, MA 01824-4105 USA
800-348-9011 Made in USA

EC|REP

ZOLL International Holding B.V.
Newtonweg 18 6662 PV ELST
The Netherlands Tel. +31 (0) 481 366410

