

**OPERATING TEMPERATURE:** 0°C to 50°C (32°F to 122°F)  
**SHORT TERM STORAGE TEMPERATURE:** -30°C to 65°C (-22°F to 149°F)  
**LONG TERM STORAGE TEMPERATURE:** 0°C to 35°C (32°F to 95°F)

STERILE R



RX ONLY 0°C to 32°F



0123

NOT MADE WITH NATURAL RUBBER LATEX.

R1345-43 Rev. L

## INDICATIONS FOR USE

*Defibrillation Noninvasive Pacing Cardioversion ECG monitoring*

For use on adult Patients with ZOLL® R Series® defibrillators by trained personnel including Physicians, Nurses, Paramedics, Emergency Medical Technicians and Cardiovascular Laboratory Technicians. The Pro-padz Sterile Adult electrodes are not indicated for use on a patient less than 8 years of age or weighing less than 55lbs (25kg).

### Instructions

1. Remove excess chest hair. Clip if necessary to maximize gel to skin contact. Clipping is recommended since shaving can leave tiny microabrasions that can lead to patient discomfort during pacing.
2. Ensure skin is clean and dry under electrode. Remove any debris, ointments, skin preps, etc. with water (and mild soap if needed). Wipe off excess moisture/diaphoresis with dry cloth.

⚠ Excessive hair can inhibit good coupling (contact), which can lead to the possibility of arcing and skin burns.


### SKIN PREPARATION

### ELECTRODE APPLICATION

#### Instructions

1. Apply one edge of the electrode securely to the patient.
2. "Roll" the electrode smoothly from that edge to the other. Be careful not to trap any pockets of air between the gel and skin.

⚠ Poor adherence and/or air under the electrodes can lead to the possibility of arcing and skin burns.



### ELECTRODE PLACEMENT


#### Anterior-Posterior (Apex/Front-Back)

Recommended for defibrillation, noninvasive pacing, ventricular cardioversion, and ECG monitoring. Optimal for noninvasive pacing because it increases patient tolerance and decreases capture thresholds.

##### Back:

Grasp the Back/Sternum electrode at the bottom and peel from the plastic liner. Place to the left of the spine just below the scapula at the heart level.

⚠ Always apply BACK electrode first. If front electrode is already in place when patient is being maneuvered for placement of the back, the front may become partially lifted. This could lead to arcing and skin burns.




BACK

##### Apex / Front:


Grasp the Apex/Front electrode at the bottom and peel from the plastic liner. Apply over cardiac apex with the nipple under adhesive area on a male patient. Position under breast on a female patient.

⚠ Avoid any contact between nipple and gel treatment area. Skin of the nipple area is more susceptible to burning.



APEX / FRONT

1. Plug BACK electrode connector into corresponding mating connector on the FRONT electrode until latched (see diagram).
2. Attach connector assembly to patient cable.



### WARNINGS

1. Contents sterile unless package opened or damaged.
2. After patient movement due to muscle contraction or patient repositioning, press pads to skin to ensure good coupling between pads and skin.
3. Do not conduct chest compressions through the pads. Doing so may cause damage to the pads that could lead to the possibility of arcing and skin burns.
4. Electrodes should be replaced after 24 hours of use or after 8 hours of continuous pacing to maximize patient benefit.
5. Do not use if gel is dry. Dried out gel can lead to skin burning. Do not open pouch until ready to use. Do not use electrodes past the expiration date printed on the pouch label.
6. To avoid electrical shock, do not touch the pads, patient, or bed when defibrillating.
7. Do not discharge standard paddles on or through electrodes or place separate ECG leads under pads. Doing so could lead to arcing and/or skin burning.
8. Always apply electrodes to flat areas of skin. If possible, avoid folds of skin such as those underneath the breast or those visible on obese individuals.
9. Avoid electrode placement near the generator of an internal pacemaker, other electrodes or metal parts in contact with the patient.
10. Some current generated by electrosurgical units (ESU) may concentrate in the conductive gel of pacing / defibrillation electrodes, especially if an ESU grounding pad other than that recommended by the ESU manufacturer is used. Consult the ESU operator's manual for further details.
11. Do not fold the electrodes or packaging. Any fold in or other damage to the conductive element could lead to the possibility of arcing and/or skin burns.
12. During prolonged pacing greater than 30 minutes, periodically examine the patient's skin for irritation.
13. Use only with ZOLL pacemaker/defibrillator products.
14. Device disposal should follow hospital protocol.

# ZOLL®

ZOLL Medical Corporation  
 269 Mill Road  
 Chelmsford, MA 01824-4105 USA  
 800-348-9011 Made in USA

EC REP ZOLL International Holding B.V.  
 Newtonweg 18 6662 PV ELST  
 The Netherlands Tel. +31 (0) 481 366410

### CARDIOVERSION

Elective cardioversion may cause visible reddening under the surface of a defibrillation / pacing / monitoring electrode. This effect is likely caused by hyperemia (excess blood) under the surface of the skin and is probably not a "burn".

During cardioversion, in contrast to a standard defibrillation, the patient is normally perfused. The impact of the energy passing through engorged capillaries under the skin's surface can cause blood to diffuse out, creating an effect that often looks like a burn or rash. The reddening typically goes away within a few days.

Among the factors that contribute to this phenomenon are:

- 1) high energy settings
- 2) multiple, successive shocks
- 3) skin integrity
- 4) patient age
- 5) certain antiarrhythmic drugs

Blistering and/or sloughing do not typically result from cardioversion and should be considered an indication of burning due to other factors.